

EnzyChrom™ Fructose Assay Kit (EFRU-100)

Quantitative Colorimetric Fructose Determination at 565nm

DESCRIPTION

FRUCTOSE (C₆H₁₂O₆, also called levulose or laevulose), is a monosaccharide found in honey, tree fruits, berries, melons, and some root vegetables along with glucose and galactose. The human body can use fructose for energy, however, too much consumption may lead to high triglycerides. Simple, direct and high-throughput assays for fructose determination find wide applications. BioAssay Systems' reagent systems reacts directly and specifically with fructose to form a colored product. Glucose and galactose do not interfere. The color intensity at 565nm is directly proportional to the fructose concentration in the sample.

KEY FEATURES

Use as little as 20 µL samples. Linear detection range in 96-well plate: 12 to 1000 µM fructose.

APPLICATIONS

Direct Assays: fructose in biological samples (e.g. serum, plasma, urine, saliva, milk, culture medium), food, juice, beverage and other agricultural products.

Drug Discovery/Pharmacology: effects of drugs on fructose metabolism.

KIT CONTENTS (100 TESTS IN 96-WELL PLATES)

Assay Buffer: 10 mL **Enzyme:** Dried
PMS Solution: 1.5 mL **Enzyme Buffer:** 150 µL
MTT Solution: 1.5 mL **Standard:** 400 µL 20 mM D-Fructose

Storage conditions. The kit is shipped on ice. Store all components at -20°C. Shelf life of six months after receipt.

Precautions: reagents are for research use only. Normal precautions for laboratory reagents should be exercised while using the reagents. Please refer to Material Safety Data Sheet for detailed information.

ASSAY PROCEDURE

Note: (1) The following substances interfere and should be avoided in sample preparation: ascorbic acid, SDS (>0.2%), sodium azide, NP-40 (>1%) and Tween-20 (>1%). (2) This assay is based on a kinetic reaction. To ensure identical incubation time, addition of Working Reagent to standard and samples should be quick and mixing should be brief but thorough. Use of a multi-channel pipettor is recommended.

Reagent Preparation: Reconstitute Enzyme by adding 120 µL Enzyme Buffer to the Enzyme tube. Make sure Enzyme is fully dissolved by pipetting up and down. Store reconstituted Enzyme at -20°C and use within 2 months.

Sample treatment: liquid samples such as serum, plasma and fruit juices can be assayed directly. Because fruit juices may contain high concentrations of fructose, it is recommended to dilute juice sample 50-fold ($n = 50$) in dH₂O prior to assay. Milk samples should be cleared by mixing 600 µL milk with 100 µL 6 N HCl. Centrifuge 5 min at 14,000 rpm. Transfer 300 µL supernatant into a clean tube and neutralize with 50 µL 6 N NaOH. The neutralized supernatant is ready for assay (dilution factor $n = 1.36$).

1. Equilibrate all components to room temperature. Briefly centrifuge the tubes before opening. Keep thawed tubes on ice during assay.
2. **Standards:** mix 12 µL 20 mM Standard with 228 µL dH₂O (final 1000 µM). Dilute standard in dH₂O as follows.

No	1000 µM STD + H ₂ O	Vol (µL)	Fructose (µM)
1	100 µL + 0 µL	100	1000
2	60 µL + 40 µL	100	600
3	30 µL + 70 µL	100	300
4	0 µL + 100 µL	100	0

Transfer 20 µL diluted standards into separate wells of a clear flat-bottom 96-well plate.

Samples: transfer 20 µL of each sample into separate wells of the plate.

3. **Color reaction.** Prepare enough Working Reagent by mixing, for each reaction well, 56 µL Assay Buffer, 1 µL Reconstituted Enzyme, 14 µL PMS Solution and 14 µL MTT Solution.

Keep Working Reagent protected from light. Add 80 µL Working Reagent to each well. Tap plate to mix. *Do not expose Working Reagent to light for more than 5 minutes.* Incubate 60 min at room temperature in the dark.

4. Read optical density at 565nm (520-600nm).

Note: If the calculated fructose concentration of a sample is higher than 1000 µM, dilute sample in water and repeat the assay. Multiply result by the dilution factor n .

CALCULATION

Subtract blank value (water, #4) from the standard values and plot the ΔOD against standard concentrations. Determine the slope and calculate the fructose concentration of Sample,


$$[\text{Fructose}] = \frac{OD_{\text{Sample}} - OD_{\text{H}_2\text{O}}}{\text{Slope } (\mu\text{M}^{-1})} \times n \quad (\mu\text{M})$$

OD_{SAMPLE} , $OD_{\text{H}_2\text{O}}$ are optical density values of the sample and water. n is the dilution factor.

Conversions: 1 mM fructose equals 18 mg/dL, 0.018% or 180 ppm.

MATERIALS REQUIRED, BUT NOT PROVIDED

Pipetting devices, centrifuge tubes, clear flat-bottom uncoated 96-well plates, optical density plate reader.



PUBLICATIONS

1. Ishimoto, Tet et al (2013). High-fat and high-sucrose (western) diet induces steatohepatitis that is dependent on fructokinase. *Hepatology*. 58(5):1632-43. Assay: Fructose in mice NA (Pubmed).
2. Li, M et al (2014). Maternal taurine supplementation attenuates maternal fructose-induced metabolic and inflammatory dysregulation and partially reverses adverse metabolic programming in offspring. *Journal of Nutritional Biochemistry* 26(3): 297-276.
3. Sharma, N (2014). Sex differences in renal and metabolic responses to a high-fructose diet in mice. *American Journal of Physiology - Renal Physiology* 308(5): F4100-10.